

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		Datamaster cdm S/N <u>130131</u> ✓	
Name <u>Shayne Calt</u>		ID# <u>4559</u> ✓	Date <u>1-21-11</u> ✓
A	Agency <u>AST</u>	Phone # <u>733-2258</u>	
Instrument Location <u>Talkeetna</u>			
Alco S/N <u>D119594</u> ✓		Target Value <u>.078</u> <u>.075</u> ^{3/8/11}	High Pressure <u>750</u>
B	Alco Test Value Average <u>.075</u> 1 st Alco		<u>.075</u> 2 nd Alco
Signature <u>[Signature]</u>			
(OVER)			

I, Nita J. Bolz, after being first duly sworn, depose and state as follows: (Do Not write in the area below)


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 14th day of MAR, 2011 ²⁰¹⁰ ₂₀₁₁

[Signature]
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office

(Notary Seal Stamp)



VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N

130131

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130131

JANUARY 21, 2011

OPERATOR'S NAME:

GILY SHAWNE H

OPERATOR'S NUMBER: 4559

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

U

O.L. #: U

DEPT/AGENCY: TALB

CASE/REPORT: U

TEST TYPE: U

ALCO TARGET VALUE: .078

ALCO S/N: D119594

— BREATH ANALYSIS —

.078 ADJUSTED FOR 29.03 in
ALCO TARGET .075 13:59
BLANK TEST .000 14:00
INTERNAL STANDARD VERIFIED 14:00
ALCO TV 29.03 in .075 14:00
BLANK TEST .000 14:01
SUBJECT SAMPLE .000 14:01
BLANK TEST .000 14:02
ALCO TV 29.03 in .075 14:03
BLANK TEST .000 14:03

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130131

JANUARY 21, 2011

TIME 13:56

— DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 43c

BREATH TUBE: 43c

BAROMETER: 29.03 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

"###&'()*+,-./0123456789:;<=>?@ABCDEF
HIJKLMNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~!"